

Technical data

VT900A + VAPOR

The Fluke Biomedical VT900A + VAPOR is designed to efficiently and reliably perform a full anesthesia machine PM, from ventilators to vaporizers.

One Solution

VT900A + VAPOR is a comprehensive test setup that can be used to test anesthesia machine ventilators and vaporizers. All flow, concentration and pressure parameters are accurately measured without any additional equipment, allowing you to lighten your load and simplify your test procedure. Calibration and servicing is made easy with one manufacturer for all anesthesia test equipment. Streamline anesthesia machine testing and improve efficiency with one test setup to meet all your needs.

Auto-detection to Ensure Patient Safety

VT900A + VAPOR automatically detects and identifies CO_2 , N_2O , sevoflurane, isoflurane, desflurane, halothane and enflurane. Two agents can be displayed simultaneously to comprehensively analyze any anesthesia flow stream making sure that only one agent is being delivered. Automatic agent identification reduces the risk of error and ensures patient safety by eliminating the need for user input and providing a means to test interlock systems and vaporizer contents. Keep patient safety the priority by testing anesthesia machines with equipment you can trust.



Key benefits and features

- Streamline your testing procedure by performing a complete anesthesia machine PM with one test setup
- Avoid confusion and improve efficiency with auto-detection of anesthetic agents and color-matched gas identification
- Reduce bulk and improve ease of transport with a convenient carrying case for the lightweight accessory and gas flow module
- Operate on-the-go with durable, robust design
- Simplify calibration and servicing with one manufacturer for all anesthesia test equipment
- Reduce testing time with 7-in. color touch screen, quick-connect fitting and complete anesthesia hosing accessory kit
- Confidently test with accuracy that meets vaporizer manufacturer recommendations
- Ensure patient safety with automatic detection and measurement of CO₂, N₂O and five commonly used anesthetic agents









Easy-to-Use

VAPOR seamlessly integrates with the VT900A test system, allowing you to quickly and easily transition from ventilator to vaporizer testing. Set up test procedures easily and read results from up to 6' (1.8m) away with the 7" (17.8 cm) touchscreen display. Reduce risk of error with auto detection and color-matched agent identification. Increase efficiency and reduce test complexity with the intuitive user interface and simple equipment setup. Anesthesia testing made easy.

Reliable Testing Anywhere

VT900A + VAPOR is a versatile, portable test solution equipped for use in a wide range of environments. Thorough drop testing and a 5-hour (minimum) battery life ensure it will continue to perform during rigorous field service. With best-in-class temperature, pressure, and humidity operating ranges and a convenient carrying case, VT900A + VAPOR can go wherever you do, continuing to perform accurate measurements that meet vaporizer manufacturer recommendations. Trust your results no matter where you test.



Weight	0.5 kg
Measured Gases	CO2, N2O, HAL, ISO, ENF, SEV, DES
Gas Corrections	Pressure and temperature
Size	191 x 96 x 57 mm
Interface	RS-232
Measurement Technology	NDIR side stream
Warmup Time ISO, full spec	45 sec / 10 minutes
Measurement Time	< 20 sec
Concentration (full accuracy*)	
CO2 % ABS range, accuracy	0-1, 0.1
	1-5, 0.2
	5-7, 0.3
	7-10, 0.5
N2O % ABS range, accuracy	0-20, 2
	20-100, 3
HAL % ABS range, accuracy	0-1, 0.15
	1-5, 0.2
SEV % ABS range, accuracy	0-1, 0.15
	1-5, 0.2
	5-8, 0.4
DES % ABS range, accuracy	0-1, 0.15
	1-5, 0.2
	5-10, 0.4
	10-15, 0.6
	15-18, 1
ISO % ABS range, accuracy	0-1, 0.15
	1-5, 0.2
ENF % ABS range, accuracy	0-1, 0.15
	1-5, 0.2
Environmental	
Operating temperature	10-40 °C
Operating humidity	10-90%

 $^{^{\}ast}$ warm-up accuracy is lower than full accuracy





Specifications

Features	
Battery life hours	8 hrs
Charge time in hours	5 hrs, typical
Memory	internal memory
Connection type	USB, Micro-B device port
Weight	3.6 lb (1.6 kg)
Display	7 in (17.8 cm)
Single full-range channel	√ (e)
Ultra-low flow ports	±750 ml/min
Ultra-low pressure port	0 to 10 mbar
Flow	o to 10 mbdi
Full range flow channel	
Range	±300 slpm
Accuracy (air)	1.7 % or 0.04 slpm
Jitra-low flow channel	1.7 % OF 0.04 SIPITI
	±750 ml/min
Range	
Accuracy (air)	±1.7 % or 0.01 slpm
Volume	1400 1
Range	±100 l ±1.75 % or 0.02 l
Accuracy	±1.75 % or 0.02 I
Pressure	
ligh pressure	
Range	-0.8 to 10 bar
ccuracy	±1 % or ±0.007 bar
Differential low pressure	1400
Range	±160 mbar
Accuracy	±0.5 % or ±0.1 mbar
JItra-low pressure	0 to 10 mb or
lange	0 to 10 mbar ±1 % or ±0.01 mbar
Accuracy	±1 % or ±0.01 mbar
virway pressure	±160 mbar
ange	±0.5 % or ±0.1 mbar
Accuracy	TO 3 % OI TO.1 IIIDAI
Barometric pressure	550 to 1240 mbar
Range Accuracy	±1 % or ±5 mbar
Other	TI /6 OI TO IIIDAI
Femperature	
Range	0 to 50 °C
Accuracy	±0.5 °C
Resolution	0.1 °C
Humidity	U.I C
Range	0 to 100 % RH
	±3 % RH (20 to 80 % RH)
Accuracy	±5 % RH (20< or >80 % RH)
Oxygen	
Range	0 to 100 %
	±1 %



Drooth parameters	
Breath parameters	0 to 60 l
Inspiratory tidal volume range	
Inspiratory tidal volume accuracy	±1.75 % or 0.5 ml
Expiratory tidal volume range	0 to 60 l
Expiratory tidal volume accuracy	±1.75 % or 0.5 ml
Minute volume range	0 to 100 l
Minute volume accuracy	±1.75 % or 0.5 ml
Breath rate range	1 to 1500 bpm
Breath rate accuracy	±1 %
Inspiratory to expiratory time ratio (I:E) range	1:300 to 300:1 +2 % or 0.1
Inspiratory to expiratory time ratio (I:E) accuracy	
Peak inspiratory pressure (PIP) range	±160 mbar +0.75 % or 0.1 mbar
Peak inspiratory pressure (PIP) accuracy	±160 mbar
Inspiratory pause pressure range	
Inspiratory pause pressure	±0.75 % or 0.1 mbar
Mean airway pressure range	±160 mbar
Mean airway pressure accuracy	±0.75 % or 0.1 mbar
Positive end expiratory pressure (PEEP) range	±160 mbar
Positive end expiratory pressure (PEEP) accuracy	±0.75 % or 0.1 mbar
Lung compliance range	0 to 1000 ml/mbar
Lung compliance accuracy	±3 % or 0.1 ml/mbar
Inspiratory time range	0 to 60 s
Inspiratory time accuracy	0.02 s
Inspiratory hold time range	0 to 60 s
Inspiratory hold time accuracy	1% or 0.1 s
Expiratory time range	0 to 90 s
Expiratory time accuracy	0.5 % or 0.01 s
Expiratory hold time range	0 to 90 s
Expiratory hold time accuracy	0.02 s
Peak expiratory flow range	±300 lpm
Peak expiratory flow accuracy	±1.7 % or 0.04 lpm
Peak inspiratory flow range	±300 lpm
Peak inspiratory flow accuracy	±1.7 % or 0.04 lpm
Environmental	10.00
Operating temp	10 °C to 40 °C
Storage temp	-20 °C to 60 °C
Operating humidity	10 to 90 % non-condensing
Storage humidity	5 to 95 % non-condensing
Gas corrections	Gas types
ATP (ambient temp/pressure, actual humidity)	Air
ATPD (ambient temp/pressure, dry)	Nitrogen (N2)
ATPS (ambient temp/pressure, saturated)	Nitrous Oxide (N2O)
STP20 (20 °C temp/pressure 760 mmHg, actual humidity)	Carbon Dioxide (CO2)
STP21 (21 °C temp/pressure 760 mmHg, actual humidity)	Oxygen (O2)
STPD0 (0 °C temp/pressure 760 mmHg, dry)	Argon
STPD20 (20 °C temp/pressure 760 mmHg, dry)	Heliox (21 % O2, 79% He)
STP or STPD21 (21 °C temp/pressure 760 mmHg, dry)	Oxygen/Nitrogen
BTPS (body temp 37 °C/ambient pressure 760 mmHg, saturated)	Oxygen/Nitrous Oxide
BTPD (body temp 37 °C/ambient pressure 760 mmHg, dry)	Oxygen/Helium



Ordering information

Includes:

- Bacterial filter (1)
- 1.2 m (4 ft) silicon tubing (2)
- 22 mm ID x 22 mm ID tubing adapters (2)
- 22 mm OD x 22 mm OD tubing adapters (2)
- Tapered 15 mm OD x 33 mm OD tubing adapters (2)
- Flexible 15 mm ID x 22 mm ID tubing adapters (2)
- DISS hand tight nut/nipple to 6.4 mm (1/4 in) ID hose barb adapter (1)
- USB serial cable
- AC power adapter
- Detachable carrying handle
- Detachable shoulder strap
- · Certificate of Calibration with test data
- VAPOR Anesthesia Tester

Optional accessories

- ACCU LUNG Test Lung
- ACCU LUNG II Test Lung
- VESA Mounting system/test arm





About Fluke Biomedical

Fluke Biomedical is the world's leading manufacturer of quality biomedical test and simulation products. In addition, Fluke Biomedical provides the latest medical imaging and oncology quality-assurance solutions for regulatory compliance. Highly credentialed and equipped with a NVLAP Lab Code 200566-0 accredited laboratory, Fluke Biomedical also offers the best in quality and customer service for all your equipment calibration needs.

Today, biomedical personnel must meet the increasing regulatory pressures, higher quality standards, and rapid technological growth, while performing their work faster and more efficiently than ever. Fluke Biomedical provides a diverse range of software and hardware tools to meet today's challenges.

Fluke Biomedical regulatory commitment

As a medical test device manufacturer, we recognize and follow certain quality standards and certifications when developing our products. We are ISO 9001 and ISO 13485 medical device certified and our products are:

- CE Certified, where required
- NIST Traceable and Calibrated
- UL, CSA, ETL Certified, where required
- NRC Compliant, where required

Fluke Biomedical.

Trusted for the measurements that matter.

Fluke Biomedical 28775 Aurora Road Cleveland, OH 44139 U.S.A.

For more information, contact us at:

(800) 850-4608 or Fax (440) 349-2307 Email: sales@flukebiomedical.com Web access: www.flukebiomedical.com

©2018 Fluke Biomedical. Specifications subject to change without notice. Printed in U.S.A. 12/2018 6011849-en

Modification of this document is not permitted without written permission from Fluke Corporation.

